

## Specifications for sputum Container

| Specifications/Requirements   | Priority   |
|---|------------|
| 1. Wide mouth   | Compulsary |
| 2. Clean  | Compulsary |
| 3. Leak proof   | Compulsary |
| 4. Screw capped   | Compulsary |
| 5. 50ml plastic transparent container with blue colour lid  | Compulsary |
| 6. Incinerable  | Compulsary |
| 7. The bidder must provide a sample   | Compulsary |
| 8. Having the NMRA registration / Authorization would Preferred   |            |
| 9. The government and NPTCCD logo should be printed on the sputum cup   |            |
| 10. Bidders could make necessary arrangement to supply 1,000,000 sputum Containers in three deliveries <ul style="list-style-type: none"> <li>• First delivery : 26<sup>th</sup> May, 2025 - 300,000 sputum Containers</li> <li>• Second delivery : 30<sup>th</sup> June, 2025 - 300,000 sputum Containers</li> <li>• Third delivery : 29<sup>th</sup> October, 2025 - 400,000 sputum Containers</li> </ul> | Compulsary |

## Sample Sputum Collection container

